

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD030419\
 Data File : VD060936.D
 Acq On : 4 Mar 2019 21:04
 Operator : VA/AP
 Sample : K1714-09
 Misc : 5.73g/5ml/MSVOA D/SOIL
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SB-B-5-8

Quant Time: Mar 05 02:56:32 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D030119S.M
 Quant Title : SW846 8260
 QLast Update : Sat Mar 02 01:28:28 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.06	168	530003	50.00	ug/l	0.02
34) 1,4-Difluorobenzene	8.23	114	905766	50.00	ug/l	0.03
63) Chlorobenzene-d5	12.37	117	802659	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	14.52	152	351749	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.46	65	229423	52.36	ug/l	0.02
Spiked Amount 50.000			Recovery	=	104.72%	
35) Dibromofluoromethane	6.93	113	361965	49.85	ug/l	0.02
Spiked Amount 50.000			Recovery	=	99.70%	
50) Toluene-d8	10.42	98	788821	43.66	ug/l	0.02
Spiked Amount 50.000			Recovery	=	87.32%	
62) 4-Bromofluorobenzene	13.56	95	331779	48.59	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.18%	
Target Compounds						
16) Acetone	3.23	43	40842	27.918	ug/l	93
20) Methylene Chloride	3.83	84	102910	12.066	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD030419\
Data File : VD060936.D
Acq On : 4 Mar 2019 21:04
Operator : VA/AP
Sample : K1714-09
Misc : 5.73g/5ml/MSVOA D/SOIL
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SB-B-5-8

Quant Time: Mar 05 02:56:32 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D030119S.M
Quant Title : SW846 8260
QLast Update : Sat Mar 02 01:28:28 2019
Response via : Initial Calibration

