

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD030419\
 Data File : VD060939.D
 Acq On : 4 Mar 2019 22:27
 Operator : VA/AP
 Sample : K1714-12
 Misc : 6.20g/5ml/MSVOA D/SOIL
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SB-B-14-18

Quant Time: Mar 05 03:04:03 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D030119S.M
 Quant Title : SW846 8260
 QLast Update : Sat Mar 02 01:28:28 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.06	168	517697	50.00	ug/l	0.02
34) 1,4-Difluorobenzene	8.23	114	896677	50.00	ug/l	0.03
63) Chlorobenzene-d5	12.38	117	835880	50.00	ug/l	0.02
72) 1,4-Dichlorobenzene-d4	14.52	152	346342	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.47	65	227209	53.09	ug/l	0.02
Spiked Amount 50.000			Recovery	=	106.18%	
35) Dibromofluoromethane	6.93	113	355997	49.52	ug/l	0.02
Spiked Amount 50.000			Recovery	=	99.04%	
50) Toluene-d8	10.42	98	820890	45.89	ug/l	0.02
Spiked Amount 50.000			Recovery	=	91.78%	
62) 4-Bromofluorobenzene	13.56	95	339123	50.17	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.34%	
Target Compounds						
16) Acetone	3.24	43	43310	30.309	ug/l	90
20) Methylene Chloride	3.84	84	124806	15.927	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD030419\
Data File : VD060939.D
Acq On : 4 Mar 2019 22:27
Operator : VA/AP
Sample : K1714-12
Misc : 6.20g/5ml/MSVOA D/SOIL
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SB-B-14-18

Quant Time: Mar 05 03:04:03 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D030119S.M
Quant Title : SW846 8260
QLast Update : Sat Mar 02 01:28:28 2019
Response via : Initial Calibration

