

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD030420\
 Data File : VD065410.D
 Acq On : 04 Mar 2020 13:47
 Operator : VA/SY
 Sample : VIBLK
 Misc : 5.00G/5.00ml/MSVOA D/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VIBLK

Quant Time: Mar 05 01:33:25 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D030420S.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 05 01:22:07 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.98	168	459712	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.87	114	726958	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.65	117	659794	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.58	152	303785	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.33	65	217034	55.04	ug/l	0.00
Spiked Amount	50.000		Recovery	=	110.08%	
35) Dibromofluoromethane	7.91	113	228856	50.91	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.82%	
50) Toluene-d8	10.34	98	822560	47.69	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.38%	
62) 4-Bromofluorobenzene	12.64	95	255225	47.94	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.88%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD030420\
Data File : VD065410.D
Acq On : 04 Mar 2020 13:47
Operator : VA/SY
Sample : VIBLK
Misc : 5.00G/5.00ml/MSVOA D/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VIBLK

Quant Time: Mar 05 01:33:25 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D030420S.M
Quant Title : SW846 8260
QLast Update : Thu Mar 05 01:22:07 2020
Response via : Initial Calibration

