

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD030421\
 Data File : VD068520.D
 Acq On : 04 Mar 2021 12:48
 Operator : VA/SY
 Sample : M1552-02
 Misc : 6.17G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 24368

Quant Time: Mar 05 00:00:37 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D022621S.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 26 14:04:14 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	357825	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.867	114	616380	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.643	117	553165	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.573	152	215812	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	181103	50.66	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	101.32%
35) Dibromofluoromethane	7.914	113	186656	50.27	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	100.54%
50) Toluene-d8	10.344	98	740233	50.05	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	100.10%
62) 4-Bromofluorobenzene	12.632	95	220086	46.17	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	92.34%
Target Compounds						
86) p-Isopropyltoluene	13.514	119	14984	1.05	ug/l	Qvalue 94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD030421\
Data File : VD068520.D
Acq On : 04 Mar 2021 12:48
Operator : VA/SY
Sample : M1552-02
Misc : 6.17G/5.00ml/MSVOA_D/SOIL
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
24368

Quant Time: Mar 05 00:00:37 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D022621S.M
Quant Title : SW846 8260
QLast Update : Fri Feb 26 14:04:14 2021
Response via : Initial Calibration

