

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD030421\
 Data File : VD068532.D
 Acq On : 04 Mar 2021 18:25
 Operator : VA/SY
 Sample : M1566-01
 Misc : 6.09G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 RO-282

Quant Time: Mar 05 00:03:43 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D022621S.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 26 14:04:14 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	340214	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.867	114	586715	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.644	117	565119	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.573	152	263530	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.338	65	170890	50.27	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	100.54%
35) Dibromofluoromethane	7.920	113	182778	51.71	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	103.42%
50) Toluene-d8	10.344	98	731747	51.98	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	103.96%
62) 4-Bromofluorobenzene	12.632	95	240650	53.04	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	106.08%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD030421\
Data File : VD068532.D
Acq On : 04 Mar 2021 18:25
Operator : VA/SY
Sample : M1566-01
Misc : 6.09G/5.00ml/MSVOA_D/SOIL
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
RO-282

Quant Time: Mar 05 00:03:43 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D022621S.M
Quant Title : SW846 8260
QLast Update : Fri Feb 26 14:04:14 2021
Response via : Initial Calibration

