

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD030421\
 Data File : VD068533.D
 Acq On : 04 Mar 2021 18:53
 Operator : VA/SY
 Sample : M1566-03
 Misc : 6.88/5.00ml/MSVOA_D/SOIL
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VEOLIA-72-11977

Quant Time: Mar 05 00:03:58 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D022621S.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 26 14:04:14 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	328663	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.867	114	573177	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.644	117	548190	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.573	152	239803	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.338	65	163421	49.77	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	99.54%
35) Dibromofluoromethane	7.920	113	175847	50.93	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	101.86%
50) Toluene-d8	10.344	98	715179	52.00	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	104.00%
62) 4-Bromofluorobenzene	12.632	95	230198	51.94	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	103.88%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD030421\
Data File : VD068533.D
Acq On : 04 Mar 2021 18:53
Operator : VA/SY
Sample : M1566-03
Misc : 6.88/5.00ml/MSVOA_D/SOIL
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VEOLIA-72-11977

Quant Time: Mar 05 00:03:58 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D022621S.M
Quant Title : SW846 8260
QLast Update : Fri Feb 26 14:04:14 2021
Response via : Initial Calibration

