

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD030421\
 Data File : VD068534.D
 Acq On : 04 Mar 2021 19:21
 Operator : VA/SY
 Sample : VSTDCCC050
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: Mar 05 00:04:14 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D022621S.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 26 14:04:14 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	80	0.00
2 T	Dichlorodifluoromethane	0.485	0.447	7.8	73	0.00
3 P	Chloromethane	0.613	0.585	4.6	78	0.00
4 C	Vinyl Chloride	0.728	0.664	8.8#	73	0.00
5 T	Bromomethane	0.518	0.493	4.8	80	0.00
6 T	Chloroethane	0.471	0.460	2.3	80	0.00
7 T	Trichlorofluoromethane	0.831	0.900	-8.3	87	0.00
8 T	Diethyl Ether	0.247	0.291	-17.8	94	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.506	0.560	-10.7	89	0.00
10 T	Methyl Iodide	0.537	0.603	-12.3	81	0.00
11 T	Tert butyl alcohol	0.027	0.030	-11.1	98	-0.01
12 CM	1,1-Dichloroethene	0.502	0.554	-10.4#	88	0.00
13 T	Acrolein	0.038	0.042	-10.5	85	0.00
14 T	Allyl chloride	0.648	0.682	-5.2	83	0.00
15 T	Acrylonitrile	0.104	0.117	-12.5	90	0.00
16 T	Acetone	0.079	0.082	-3.8	81	0.00
17 T	Carbon Disulfide	1.645	1.753	-6.6	85	0.00
18 T	Methyl Acetate	0.205	0.245	-19.5	100	0.00
19 T	Methyl tert-butyl Ether	1.089	1.147	-5.3	86	-0.01
20 T	Methylene Chloride	0.673	0.738	-9.7	98	0.00
21 T	trans-1,2-Dichloroethene	0.600	0.595	0.8	80	0.00
22 T	Diisopropyl ether	1.428	1.406	1.5	78	0.00
23 T	Vinyl Acetate	0.741	0.752	-1.5	80	0.00
24 P	1,1-Dichloroethane	0.979	0.977	0.2	80	0.00
25 T	2-Butanone	0.114	0.114	0.0	81	0.00
26 T	2,2-Dichloropropane	0.815	0.762	6.5	75	0.00
27 T	cis-1,2-Dichloroethene	0.658	0.651	1.1	79	0.00
28 T	Bromochloromethane	0.382	0.391	-2.4	81	0.00
29 T	Tetrahydrofuran	0.074	0.074	0.0	80	0.00
30 C	Chloroform	1.033	1.078	-4.4#	84	0.00
31 T	Cyclohexane	0.909	0.775	14.7	71	0.00
32 T	1,1,1-Trichloroethane	0.882	0.877	0.6	80	0.00
33 S	1,2-Dichloroethane-d4	0.500	0.504	-0.8	81	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	80	0.00
35 S	Dibromofluoromethane	0.301	0.312	-3.7	84	0.00
36 T	1,1-Dichloropropene	0.475	0.442	6.9	75	0.00
37 T	Ethyl Acetate	0.158	0.157	0.6	83	0.00
38 T	Carbon Tetrachloride	0.447	0.434	2.9	79	0.00
39 T	Methylcyclohexane	0.586	0.537	8.4	72	0.00
40 TM	Benzene	1.363	1.344	1.4	80	0.00
41 T	Methacrylonitrile	0.092	0.089	3.3	84	0.00
42 TM	1,2-Dichloroethane	0.358	0.361	-0.8	81	0.00
43 T	Isopropyl Acetate	0.290	0.304	-4.8	84	0.00
44 TM	Trichloroethene	0.374	0.378	-1.1	82	0.00
45 C	1,2-Dichloropropane	0.326	0.329	-0.9#	80	0.00
46 T	Dibromomethane	0.177	0.186	-5.1	87	0.00
47 T	Bromodichloromethane	0.459	0.492	-7.2	87	0.00
48 T	Methyl methacrylate	0.154	0.146	5.2	77	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD030421\
 Data File : VD068534.D
 Acq On : 04 Mar 2021 19:21
 Operator : VA/SY
 Sample : VSTDCCC050
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: Mar 05 00:04:14 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D022621S.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 26 14:04:14 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.002	0.002	0.0	86	0.00
50 S	Toluene-d8	1.200	1.159	3.4	79	0.00
51 T	4-Methyl-2-Pentanone	0.152	0.159	-4.6	83	0.00
52 CM	Toluene	0.871	0.871	0.0	81	0.00
53 T	t-1,3-Dichloropropene	0.424	0.442	-4.2	85	0.00
54 T	cis-1,3-Dichloropropene	0.517	0.514	0.6	80	0.00
55 T	1,1,2-Trichloroethane	0.245	0.264	-7.8	87	0.00
56 T	Ethyl methacrylate	0.282	0.298	-5.7	83	0.00
57 T	1,3-Dichloropropane	0.428	0.441	-3.0	84	0.00
58 T	2-Chloroethyl Vinyl ether	0.141	0.147	-4.3	83	0.00
59 T	2-Hexanone	0.101	0.106	-5.0	82	0.00
60 T	Dibromochloromethane	0.304	0.332	-9.2	87	0.00
61 T	1,2-Dibromoethane	0.240	0.255	-6.3	86	0.00
62 S	4-Bromofluorobenzene	0.387	0.389	-0.5	81	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	81	0.00
64 T	Tetrachloroethene	0.327	0.324	0.9	83	0.00
65 PM	Chlorobenzene	1.018	1.014	0.4	82	0.00
66 T	1,1,1,2-Tetrachloroethane	0.354	0.373	-5.4	87	0.00
67 C	Ethyl Benzene	1.818	1.790	1.5#	80	0.00
68 T	m/p-Xylenes	0.690	0.692	-0.3	81	0.00
69 T	o-Xylene	0.636	0.644	-1.3	81	0.00
70 T	Styrene	1.087	1.140	-4.9	83	0.00
71 P	Bromoform	0.180	0.199	-10.6	89	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	87	0.00
73 T	Isopropylbenzene	3.708	3.406	8.1	80	0.00
74 T	N-amyl acetate	0.632	0.630	0.3	86	0.00
75 P	1,1,2,2-Tetrachloroethane	0.638	0.605	5.2	86	0.00
76 T	1,2,3-Trichloropropane	0.444	0.428	3.6	100	0.00
77 T	Bromobenzene	0.846	0.823	2.7	87	0.00
78 T	n-propylbenzene	4.431	4.076	8.0	80	0.00
79 T	2-Chlorotoluene	2.459	2.277	7.4	81	0.00
80 T	1,3,5-Trimethylbenzene	3.039	2.864	5.8	81	0.00
81 T	trans-1,4-Dichloro-2-butene	0.190	0.183	3.7	83	0.00
82 T	4-Chlorotoluene	2.588	2.414	6.7	81	0.00
83 T	tert-Butylbenzene	2.588	2.431	6.1	81	0.00
84 T	1,2,4-Trimethylbenzene	3.044	2.928	3.8	82	0.00
85 T	sec-Butylbenzene	3.656	3.369	7.9	80	0.00
86 T	p-Isopropyltoluene	3.318	3.140	5.4	81	0.00
87 T	1,3-Dichlorobenzene	1.635	1.598	2.3	85	0.00
88 T	1,4-Dichlorobenzene	1.619	1.583	2.2	87	0.00
89 T	n-Butylbenzene	3.201	2.900	9.4	77	0.00
90 T	Hexachloroethane	0.660	0.693	-5.0	88	0.00
91 T	1,2-Dichlorobenzene	1.402	1.407	-0.4	88	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.112	0.105	6.3	85	0.00
93 T	1,2,4-Trichlorobenzene	1.083	1.082	0.1	86	0.00
94 T	Hexachlorobutadiene	0.601	0.579	3.7	83	0.00
95 T	Naphthalene	1.961	1.975	-0.7	87	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD030421\
Data File : VD068534.D
Acq On : 04 Mar 2021 19:21
Operator : VA/SY
Sample : VSTDCCC050
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_D
LabSampleId :
VSTDCCC050

Quant Time: Mar 05 00:04:14 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D022621S.M
Quant Title : SW846 8260
QLast Update : Fri Feb 26 14:04:14 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.937	0.933	0.4	87	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 5