

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD030424\
 Data File : VD078594.D
 Acq On : 04 Mar 2024 18:31
 Operator : RP/MD
 Sample : P1669-11
 Misc : 4.88G/5ml/MSVOA_D/SOIL/A
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 TP-6-VOC

Quant Time: Mar 05 04:16:57 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D030124S.M
 Quant Title : SW846 8260
 QLast Update : Sat Mar 02 03:50:17 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	122869	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	238158	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	219527	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	87695	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.234	65	76247	62.601	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	125.200%
35) Dibromofluoromethane	7.804	113	85165	54.973	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	109.940%
50) Toluene-d8	10.269	98	302884	50.314	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	100.620%
62) 4-Bromofluorobenzene	12.575	95	77265	44.937	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	89.880%
Target Compounds						
						Qvalue
16) Acetone	4.022	43	13132	56.277	ug/l	96
25) 2-Butanone	7.087	43	2438	8.283	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD030424\
Data File : VD078594.D
Acq On : 04 Mar 2024 18:31
Operator : RP/MD
Sample : P1669-11
Misc : 4.88G/5ml/MSVOA_D/SOIL/A
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
TP-6-VOC

Quant Time: Mar 05 04:16:57 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D030124S.M
Quant Title : SW846 8260
QLast Update : Sat Mar 02 03:50:17 2024
Response via : Initial Calibration

