

Data Path : Z:\voasrv\HPCHEM1\MSVOA D\Data\VD030520\
 Data File : VD065439.D
 Acq On : 05 Mar 2020 18:59
 Operator : VA/SY
 Sample : VSTDCCC050
 Misc : 5.00G/5.00ml/MSVOA D/SOIL
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: Mar 06 04:16:09 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D030420S.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 05 01:22:07 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	89	0.00
2 T	Dichlorodifluoromethane	0.430	0.371	13.7	84	0.00
3 P	Chloromethane	0.519	0.472	9.1	89	0.00
4 C	Vinyl Chloride	0.512	0.476	7.0#	86	0.00
5 T	Bromomethane	0.327	0.311	4.9	91	0.00
6 T	Chloroethane	0.313	0.299	4.5	87	0.00
7 T	Trichlorofluoromethane	0.754	0.712	5.6	88	0.00
8 T	Diethyl Ether	0.224	0.230	-2.7	96	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.486	0.457	6.0	89	0.00
10 T	Methyl Iodide	0.505	0.518	-2.6	84	0.00
11 T	Tert butyl alcohol	0.025	0.022	12.0	88	0.00
12 CM	1,1-Dichloroethene	0.465	0.434	6.7#	87	0.00
13 T	Acrolein	0.042	0.041	2.4	96	0.00
14 T	Allyl chloride	0.714	0.704	1.4	89	0.00
15 T	Acrylonitrile	0.098	0.104	-6.1	99	0.00
16 T	Acetone	0.087	0.081	6.9	91	0.00
17 T	Carbon Disulfide	1.549	1.449	6.5	86	0.00
18 T	Methyl Acetate	0.220	0.227	-3.2	99	0.00
19 T	Methyl tert-butyl Ether	0.894	0.977	-9.3	97	0.00
20 T	Methylene Chloride	0.537	0.509	5.2	95	0.00
21 T	trans-1,2-Dichloroethene	0.517	0.515	0.4	90	0.00
22 T	Diisopropyl ether	1.328	1.460	-9.9	96	0.00
23 T	Vinyl Acetate	0.737	0.833	-13.0	97	0.00
24 P	1,1-Dichloroethane	0.852	0.862	-1.2	92	0.00
25 T	2-Butanone	0.118	0.123	-4.2	97	0.00
26 T	2,2-Dichloropropane	0.744	0.712	4.3	88	0.00
27 T	cis-1,2-Dichloroethene	0.529	0.545	-3.0	92	0.00
28 T	Bromochloromethane	0.341	0.370	-8.5	100	0.00
29 T	Tetrahydrofuran	0.076	0.084	-10.5	100	0.00
30 C	Chloroform	0.851	0.884	-3.9#	94	0.00
31 T	Cyclohexane	0.852	0.798	6.3	87	0.00
32 T	1,1,1-Trichloroethane	0.787	0.774	1.7	89	0.00
33 S	1,2-Dichloroethane-d4	0.429	0.442	-3.0	95	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	93	0.00
35 S	Dibromofluoromethane	0.309	0.311	-0.6	93	0.00
36 T	1,1-Dichloropropene	0.479	0.474	1.0	90	0.00
37 T	Ethyl Acetate	0.185	0.200	-8.1	103	0.00
38 T	Carbon Tetrachloride	0.497	0.487	2.0	91	0.00
39 T	Methylcyclohexane	0.579	0.563	2.8	86	0.00
40 TM	Benzene	1.340	1.351	-0.8	92	0.00
41 T	Methacrylonitrile	0.098	0.113	-15.3	123	0.01
42 TM	1,2-Dichloroethane	0.350	0.367	-4.9	99	0.00
43 T	Isopropyl Acetate	0.342	0.355	-3.8	97	0.00
44 TM	Trichloroethene	0.378	0.374	1.1	91	0.00
45 C	1,2-Dichloropropane	0.325	0.336	-3.4#	95	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA D\Data\VD030520\
 Data File : VD065439.D
 Acq On : 05 Mar 2020 18:59
 Operator : VA/SY
 Sample : VSTDCCC050
 Misc : 5.00G/5.00ml/MSVOA D/SOIL
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: Mar 06 04:16:09 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D030420S.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 05 01:22:07 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.168	0.176	-4.8	96	0.00
47 T	Bromodichloromethane	0.438	0.457	-4.3	96	0.00
48 T	Methyl methacrylate	0.161	0.171	-6.2	99	0.00
49 T	1,4-Dioxane	0.002	0.002	0.0	98	0.00
50 S	Toluene-d8	1.186	1.208	-1.9	90	0.00
51 T	4-Methyl-2-Pentanone	0.172	0.193	-12.2	102	0.00
52 CM	Toluene	0.844	0.866	-2.6#	92	0.00
53 T	t-1,3-Dichloropropene	0.409	0.437	-6.8	97	0.00
54 T	cis-1,3-Dichloropropene	0.495	0.525	-6.1	95	0.00
55 T	1,1,2-Trichloroethane	0.238	0.247	-3.8	97	0.00
56 T	Ethyl methacrylate	0.269	0.300	-11.5	99	0.00
57 T	1,3-Dichloropropane	0.403	0.427	-6.0	98	0.00
58 T	2-Chloroethyl Vinyl ether	0.087	0.083	4.6	100	0.00
59 T	2-Hexanone	0.119	0.132	-10.9	99	0.00
60 T	Dibromochloromethane	0.306	0.323	-5.6	98	0.00
61 T	1,2-Dibromoethane	0.227	0.241	-6.2	99	0.00
62 S	4-Bromofluorobenzene	0.366	0.382	-4.4	92	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	93	0.00
64 T	Tetrachloroethene	0.358	0.351	2.0	91	0.00
65 PM	Chlorobenzene	0.989	1.011	-2.2	95	0.00
66 T	1,1,1,2-Tetrachloroethane	0.360	0.376	-4.4	97	0.00
67 C	Ethyl Benzene	1.710	1.769	-3.5#	92	0.00
68 T	m/p-Xylenes	0.659	0.691	-4.9	92	0.00
69 T	o-Xylene	0.580	0.618	-6.6	93	0.00
70 T	Styrene	1.021	1.107	-8.4	95	0.00
71 P	Bromoform	0.202	0.212	-5.0	100	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	97	0.00
73 T	Isopropylbenzene	3.289	3.336	-1.4	91	0.00
74 T	N-amyl acetate	0.673	0.737	-9.5	103	0.00
75 P	1,1,2,2-Tetrachloroethane	0.578	0.604	-4.5	103	0.00
76 T	1,2,3-Trichloropropane	0.395	0.447	-13.2	112	0.00
77 T	Bromobenzene	0.817	0.843	-3.2	98	0.00
78 T	n-propylbenzene	3.902	4.000	-2.5	93	0.00
79 T	2-Chlorotoluene	2.199	2.215	-0.7	94	0.00
80 T	1,3,5-Trimethylbenzene	2.704	2.769	-2.4	92	0.00
81 T	trans-1,4-Dichloro-2-butene	0.187	0.190	-1.6	98	0.00
82 T	4-Chlorotoluene	2.311	2.366	-2.4	95	0.00
83 T	tert-Butylbenzene	2.287	2.331	-1.9	93	0.00
84 T	1,2,4-Trimethylbenzene	2.665	2.776	-4.2	93	0.00
85 T	sec-Butylbenzene	3.265	3.280	-0.5	90	0.00
86 T	p-Isopropyltoluene	2.979	3.050	-2.4	91	0.00
87 T	1,3-Dichlorobenzene	1.571	1.565	0.4	96	0.00
88 T	1,4-Dichlorobenzene	1.563	1.534	1.9	95	0.00
89 T	n-Butylbenzene	2.754	2.778	-0.9	90	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA D\Data\VD030520\
Data File : VD065439.D
Acq On : 05 Mar 2020 18:59
Operator : VA/SY
Sample : VSTDCCC050
Misc : 5.00G/5.00ml/MSVOA D/SOIL
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_D
LabSampleId :
VSTDCCC050

Quant Time: Mar 06 04:16:09 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D030420S.M
Quant Title : SW846 8260
QLast Update : Thu Mar 05 01:22:07 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.614	0.602	2.0	93	0.00
91 T	1,2-Dichlorobenzene	1.337	1.376	-2.9	99	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.088	0.090	-2.3	102	0.00
93 T	1,2,4-Trichlorobenzene	0.906	0.934	-3.1	98	0.00
94 T	Hexachlorobutadiene	0.595	0.574	3.5	92	0.00
95 T	Naphthalene	1.449	1.579	-9.0	101	0.00
96 T	1,2,3-Trichlorobenzene	0.783	0.816	-4.2	98	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6