

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD030521\
 Data File : VD068548.D
 Acq On : 05 Mar 2021 17:21
 Operator : VA/SY
 Sample : M1582-04
 Misc : 5.15G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 RO-363

Quant Time: Mar 05 23:13:40 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D022621S.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 26 14:04:14 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.985	168	21270	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	33942	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.649	117	33121	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.573	152	16075	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	13171	61.98	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 123.96%	
35) Dibromofluoromethane	7.920	113	9451	46.22	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 92.44%	
50) Toluene-d8	10.344	98	39400	48.38	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 96.76%	
62) 4-Bromofluorobenzene	12.632	95	14142	53.88	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 107.76%	
Target Compounds						
84) 1,2,4-Trimethylbenzene	13.273	105	1594	1.63	ug/l	Qvalue 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD030521\
Data File : VD068548.D
Acq On : 05 Mar 2021 17:21
Operator : VA/SY
Sample : M1582-04
Misc : 5.15G/5.00ml/MSVOA_D/SOIL
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
RO-363

Quant Time: Mar 05 23:13:40 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D022621S.M
Quant Title : SW846 8260
QLast Update : Fri Feb 26 14:04:14 2021
Response via : Initial Calibration

