

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD030521\
 Data File : VD068549.D
 Acq On : 05 Mar 2021 17:49
 Operator : VA/SY
 Sample : M1582-08
 Misc : 4.55G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VEOLIA-72-11991

Quant Time: Mar 05 23:13:55 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D022621S.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 26 14:04:14 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.985	168	8516	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.873	114	13524	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.644	117	12318	50.00	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	6019	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	4979	58.52	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	117.04%
35) Dibromofluoromethane	7.914	113	3714	45.59	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	91.18%
50) Toluene-d8	10.338	98	15768	48.59	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	97.18%
62) 4-Bromofluorobenzene	12.632	95	5035	48.14	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	96.28%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD030521\
Data File : VD068549.D
Acq On : 05 Mar 2021 17:49
Operator : VA/SY
Sample : M1582-08
Misc : 4.55G/5.00ml/MSVOA_D/SOIL
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VEOLIA-72-11991

Quant Time: Mar 05 23:13:55 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D022621S.M
Quant Title : SW846 8260
QLast Update : Fri Feb 26 14:04:14 2021
Response via : Initial Calibration

