



Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD030521\
Data File : VD068535.D
Acq On : 05 Mar 2021 09:44
Operator : VA/SY
Sample : BFB
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D022621S.M
Title : SW846 8260
Last Update : Fri Feb 26 14:04:14 2021



AutoFind: Scans 1820, 1821, 1822; Background Corrected with Scan 1812

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	18.5	13630	PASS
75	95	30	60	51.6	38051	PASS
95	95	100	100	100.0	73701	PASS
96	95	5	9	6.2	4543	PASS
173	174	0.00	2	1.1	739	PASS
174	95	50	100	90.7	66845	PASS
175	174	5	9	8.0	5368	PASS
176	174	95	101	95.1	63603	PASS
177	176	5	9	6.7	4276	PASS