

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD030819\
 Data File : VD060977.D
 Acq On : 8 Mar 2019 11:53
 Operator : VA/AP
 Sample : VD0308SBL01
 Misc : 5.00g/5ml/MSVOA D/SOIL
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VD0308SBL01

Quant Time: Mar 08 23:35:56 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D030719S.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 08 02:35:39 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.03	168	903113	50.00	ug/l	-0.01
34) 1,4-Difluorobenzene	8.20	114	1707796	50.00	ug/l	-0.01
63) Chlorobenzene-d5	12.36	117	1255227	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	14.51	152	561877	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	7.45	65	442580	58.87	ug/l	0.00
Spiked Amount						
			Recovery	=	117.74%	
35) Dibromofluoromethane	6.91	113	678338	53.52	ug/l	-0.01
Spiked Amount						
			Recovery	=	107.04%	
50) Toluene-d8	10.40	98	1610455	50.60	ug/l	0.00
Spiked Amount						
			Recovery	=	101.20%	
62) 4-Bromofluorobenzene	13.55	95	649429	52.59	ug/l	0.00
Spiked Amount						
			Recovery	=	105.18%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD030819\
Data File : VD060977.D
Acq On : 8 Mar 2019 11:53
Operator : VA/AP
Sample : VD0308SBL01
Misc : 5.00g/5ml/MSVOA D/SOIL
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VD0308SBL01

Quant Time: Mar 08 23:35:56 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D030719S.M
Quant Title : SW846 8260
QLast Update : Fri Mar 08 02:35:39 2019
Response via : Initial Calibration

