



Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD030824\
Data File : VD078612.D
Acq On : 08 Mar 2024 09:42
Operator : RP/MD
Sample : BFB
Misc : 5.00G/5ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D030124S.M
Title : SW846 8260
Last Update : Sat Mar 02 03:50:17 2024



AutoFind: Scans 1867, 1868, 1869; Background Corrected with Scan 1859

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
50	95	15	40	21.2	6825	PASS
75	95	30	60	55.7	17968	PASS
95	95	100	100	100.0	32237	PASS
96	95	5	9	7.0	2271	PASS
173	174	0.00	2	1.2	384	PASS
174	95	50	100	97.0	31267	PASS
175	174	5	9	7.8	2445	PASS
176	174	95	101	98.7	30859	PASS
177	176	5	9	6.3	1937	PASS