

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD030921\
 Data File : VD068573.D
 Acq On : 09 Mar 2021 12:19
 Operator : VA/SY
 Sample : M1599-01
 Misc : 5.34G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 OK-02-030821

Quant Time: Mar 10 00:00:35 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D022621S.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 26 14:04:14 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.985	168	326961	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.867	114	590361	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.643	117	558673	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.573	152	241469	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	179497	54.95	ug/l	0.00
Spiked Amount 50.000	Range 50 - 163		Recovery = 109.90%			
35) Dibromofluoromethane	7.920	113	183431	51.58	ug/l	0.00
Spiked Amount 50.000	Range 54 - 147		Recovery = 103.16%			
50) Toluene-d8	10.343	98	735116	51.90	ug/l	0.00
Spiked Amount 50.000	Range 49 - 140		Recovery = 103.80%			
62) 4-Bromofluorobenzene	12.631	95	235620	51.61	ug/l	0.00
Spiked Amount 50.000	Range 25 - 144		Recovery = 103.22%			
Target Compounds						Qvalue
16) Acetone	4.167	43	25574	49.28	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD030921\
Data File : VD068573.D
Acq On : 09 Mar 2021 12:19
Operator : VA/SY
Sample : M1599-01
Misc : 5.34G/5.00ml/MSVOA_D/SOIL
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
OK-02-030821

Quant Time: Mar 10 00:00:35 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D022621S.M
Quant Title : SW846 8260
QLast Update : Fri Feb 26 14:04:14 2021
Response via : Initial Calibration

