

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD030921\
 Data File : VD068576.D
 Acq On : 09 Mar 2021 15:09
 Operator : VA/SY
 Sample : VSTDCCC050
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: Mar 10 00:01:16 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D022621S.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 26 14:04:14 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	84	0.00
2 T	Dichlorodifluoromethane	0.485	0.439	9.5	75	0.00
3 P	Chloromethane	0.613	0.561	8.5	78	0.00
4 C	Vinyl Chloride	0.728	0.683	6.2#	78	0.00
5 T	Bromomethane	0.518	0.490	5.4	84	0.00
6 T	Chloroethane	0.471	0.452	4.0	82	0.00
7 T	Trichlorofluoromethane	0.831	0.810	2.5	82	0.00
8 T	Diethyl Ether	0.247	0.246	0.4	83	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.506	0.495	2.2	83	0.00
10 T	Methyl Iodide	0.537	0.583	-8.6	82	0.00
11 T	Tert butyl alcohol	0.027	0.030	-11.1	104	0.00
12 CM	1,1-Dichloroethene	0.502	0.497	1.0#	83	0.00
13 T	Acrolein	0.038	0.029	23.7	62	0.00
14 T	Allyl chloride	0.648	0.654	-0.9	84	0.00
15 T	Acrylonitrile	0.104	0.123	-18.3	99	0.00
16 T	Acetone	0.079	0.072	8.9	74	0.00
17 T	Carbon Disulfide	1.645	1.528	7.1	77	0.00
18 T	Methyl Acetate	0.205	0.234	-14.1	100	0.00
19 T	Methyl tert-butyl Ether	1.089	1.305	-19.8	102	0.00
20 T	Methylene Chloride	0.673	0.652	3.1	90	0.00
21 T	trans-1,2-Dichloroethene	0.600	0.691	-15.2	97	0.00
22 T	Diisopropyl ether	1.428	1.460	-2.2	85	0.00
23 T	Vinyl Acetate	0.741	0.769	-3.8	86	0.00
24 P	1,1-Dichloroethane	0.979	0.996	-1.7	85	-0.01
25 T	2-Butanone	0.114	0.114	0.0	85	0.00
26 T	2,2-Dichloropropane	0.815	0.833	-2.2	86	0.00
27 T	cis-1,2-Dichloroethene	0.658	0.668	-1.5	85	0.00
28 T	Bromochloromethane	0.382	0.390	-2.1	85	0.00
29 T	Tetrahydrofuran	0.074	0.077	-4.1	88	0.00
30 C	Chloroform	1.033	1.043	-1.0#	85	0.00
31 T	Cyclohexane	0.909	0.833	8.4	80	0.00
32 T	1,1,1-Trichloroethane	0.882	0.897	-1.7	85	0.00
33 S	1,2-Dichloroethane-d4	0.500	0.518	-3.6	88	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	83	0.00
35 S	Dibromofluoromethane	0.301	0.322	-7.0	90	0.00
36 T	1,1-Dichloropropene	0.475	0.461	2.9	81	0.00
37 T	Ethyl Acetate	0.158	0.154	2.5	85	0.00
38 T	Carbon Tetrachloride	0.447	0.456	-2.0	86	0.00
39 T	Methylcyclohexane	0.586	0.572	2.4	80	0.00
40 TM	Benzene	1.363	1.381	-1.3	85	0.00
41 T	Methacrylonitrile	0.092	0.092	0.0	90	0.00
42 TM	1,2-Dichloroethane	0.358	0.362	-1.1	84	0.00
43 T	Isopropyl Acetate	0.290	0.311	-7.2	88	0.00
44 TM	Trichloroethene	0.374	0.376	-0.5	85	0.00
45 C	1,2-Dichloropropane	0.326	0.336	-3.1#	85	0.00
46 T	Dibromomethane	0.177	0.184	-4.0	90	0.00
47 T	Bromodichloromethane	0.459	0.480	-4.6	87	0.00
48 T	Methyl methacrylate	0.154	0.157	-1.9	86	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD030921\
 Data File : VD068576.D
 Acq On : 09 Mar 2021 15:09
 Operator : VA/SY
 Sample : VSTDCCC050
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: Mar 10 00:01:16 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D022621S.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 26 14:04:14 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.002	0.002	0.0	92	0.00
50 S	Toluene-d8	1.200	1.210	-0.8	85	0.00
51 T	4-Methyl-2-Pentanone	0.152	0.162	-6.6	88	0.00
52 CM	Toluene	0.871	0.899	-3.2#	86	0.00
53 T	t-1,3-Dichloropropene	0.424	0.450	-6.1	90	0.00
54 T	cis-1,3-Dichloropropene	0.517	0.522	-1.0	84	0.00
55 T	1,1,2-Trichloroethane	0.245	0.263	-7.3	89	0.00
56 T	Ethyl methacrylate	0.282	0.310	-9.9	90	0.00
57 T	1,3-Dichloropropane	0.428	0.444	-3.7	87	0.00
58 T	2-Chloroethyl Vinyl ether	0.141	0.150	-6.4	88	0.00
59 T	2-Hexanone	0.101	0.108	-6.9	87	0.00
60 T	Dibromochloromethane	0.304	0.329	-8.2	89	0.00
61 T	1,2-Dibromoethane	0.240	0.249	-3.8	87	0.00
62 S	4-Bromofluorobenzene	0.387	0.407	-5.2	88	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	84	0.00
64 T	Tetrachloroethene	0.327	0.325	0.6	86	0.00
65 PM	Chlorobenzene	1.018	1.037	-1.9	87	0.00
66 T	1,1,1,2-Tetrachloroethane	0.354	0.375	-5.9	90	0.00
67 C	Ethyl Benzene	1.818	1.856	-2.1#	85	0.00
68 T	m/p-Xylenes	0.690	0.708	-2.6	85	0.00
69 T	o-Xylene	0.636	0.660	-3.8	85	0.00
70 T	Styrene	1.087	1.160	-6.7	88	0.00
71 P	Bromoform	0.180	0.201	-11.7	93	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	85	0.00
73 T	Isopropylbenzene	3.708	3.712	-0.1	85	0.00
74 T	N-amyl acetate	0.632	0.659	-4.3	87	0.00
75 P	1,1,2,2-Tetrachloroethane	0.638	0.648	-1.6	90	0.00
76 T	1,2,3-Trichloropropane	0.444	0.436	1.8	99	0.00
77 T	Bromobenzene	0.846	0.863	-2.0	89	0.00
78 T	n-propylbenzene	4.431	4.435	-0.1	85	0.00
79 T	2-Chlorotoluene	2.459	2.502	-1.7	87	0.00
80 T	1,3,5-Trimethylbenzene	3.039	3.118	-2.6	86	0.00
81 T	trans-1,4-Dichloro-2-butene	0.190	0.196	-3.2	86	0.00
82 T	4-Chlorotoluene	2.588	2.598	-0.4	85	0.00
83 T	tert-Butylbenzene	2.588	2.631	-1.7	85	0.00
84 T	1,2,4-Trimethylbenzene	3.044	3.146	-3.4	86	0.00
85 T	sec-Butylbenzene	3.656	3.769	-3.1	87	0.00
86 T	p-Isopropyltoluene	3.318	3.440	-3.7	86	0.00
87 T	1,3-Dichlorobenzene	1.635	1.709	-4.5	88	0.00
88 T	1,4-Dichlorobenzene	1.619	1.671	-3.2	89	0.00
89 T	n-Butylbenzene	3.201	3.234	-1.0	84	0.00
90 T	Hexachloroethane	0.660	0.708	-7.3	87	0.00
91 T	1,2-Dichlorobenzene	1.402	1.463	-4.4	90	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.112	0.115	-2.7	91	0.00
93 T	1,2,4-Trichlorobenzene	1.083	1.142	-5.4	88	0.00
94 T	Hexachlorobutadiene	0.601	0.633	-5.3	88	0.00
95 T	Naphthalene	1.961	2.085	-6.3	89	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD030921\
Data File : VD068576.D
Acq On : 09 Mar 2021 15:09
Operator : VA/SY
Sample : VSTDCCC050
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_D
LabSampleId :
VSTDCCC050

Quant Time: Mar 10 00:01:16 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D022621S.M
Quant Title : SW846 8260
QLast Update : Fri Feb 26 14:04:14 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.937	0.976	-4.2	89	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6