


Data Path : Z:\voasrv\HPCHEM1\MSVOA D\Data\VD031020\
Data File : VD065458.D
Acq On : 10 Mar 2020 10:22
Operator : VA/SY
Sample : BFB
Misc : 5.00G/5.00ml/MSVOA D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D030420S.M
Title : SW846 8260
Last Update : Thu Mar 05 01:22:07 2020



AutoFind: Scans 1822, 1823, 1824; Background Corrected with Scan 1813

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	17.8	52320	PASS
75	95	30	60	48.4	141850	PASS
95	95	100	100	100.0	293226	PASS
96	95	5	9	6.4	18893	PASS
173	174	0.00	2	1.4	3658	PASS
174	95	50	100	92.0	269760	PASS
175	174	5	9	7.5	20240	PASS
176	174	95	101	97.8	263744	PASS
177	176	5	9	6.7	17576	PASS