

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD031120\
 Data File : VD065478.D
 Acq On : 11 Mar 2020 13:34
 Operator : VA/SY
 Sample : L1864-04RE
 Misc : 7.67G/5.00ml/MSVOA D/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SP-2RE

Quant Time: Mar 11 16:41:24 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D030420S.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 05 01:22:07 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.98	168	519194	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.87	114	889969	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.65	117	831071	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.58	152	381768	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.34	65	274703	61.68	ug/l	0.00
Spiked Amount 50.000			Recovery	=	123.36%	
35) Dibromofluoromethane	7.92	113	291881	53.03	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.06%	
50) Toluene-d8	10.34	98	1077329	51.02	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.04%	
62) 4-Bromofluorobenzene	12.64	95	336636	51.65	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.30%	
Target Compounds						
16) Acetone	4.17	43	24399	27.089	ug/l	95
20) Methylene Chloride	4.96	84	27312	2.286	ug/l	96
70) Styrene	12.20	104	22711	1.339	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD031120\
Data File : VD065478.D
Acq On : 11 Mar 2020 13:34
Operator : VA/SY
Sample : L1864-04RE
Misc : 7.67G/5.00ml/MSVOA D/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SP-2RE

Quant Time: Mar 11 16:41:24 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D030420S.M
Quant Title : SW846 8260
QLast Update : Thu Mar 05 01:22:07 2020
Response via : Initial Calibration

