

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD031121\
 Data File : VD068579.D
 Acq On : 11 Mar 2021 10:46
 Operator : VA/SY
 Sample : VD0311SBL01
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VD0311SBL01

Quant Time: Mar 12 00:09:37 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D022621S.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 26 14:04:14 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	363201	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.867	114	650840	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.644	117	626197	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.573	152	289707	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	184420	50.82	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	101.64%
35) Dibromofluoromethane	7.920	113	199561	50.90	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	101.80%
50) Toluene-d8	10.344	98	816190	52.26	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	104.52%
62) 4-Bromofluorobenzene	12.632	95	274747	54.59	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	109.18%

Target Compounds	Qvalue
-----	-----

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD031121\
Data File : VD068579.D
Acq On : 11 Mar 2021 10:46
Operator : VA/SY
Sample : VD0311SBL01
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VD0311SBL01

Quant Time: Mar 12 00:09:37 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D022621S.M
Quant Title : SW846 8260
QLast Update : Fri Feb 26 14:04:14 2021
Response via : Initial Calibration

