

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD031121\
 Data File : VD068592.D
 Acq On : 11 Mar 2021 18:30
 Operator : VA/SY
 Sample : M1642-03
 Misc : 6.36G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 CSB-6-8-8.5

Quant Time: Mar 12 00:14:33 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D022621S.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 26 14:04:14 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	496554	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.867	114	872935	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.644	117	854494	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.573	152	492445	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	246735	49.73	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	99.46%
35) Dibromofluoromethane	7.920	113	261900	49.80	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	99.60%
50) Toluene-d8	10.338	98	1086032	51.85	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	103.70%
62) 4-Bromofluorobenzene	12.626	95	438448	64.95	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	129.90%
Target Compounds						
16) Acetone	4.162	43	7408	9.40	ug/l	95
39) Methylcyclohexane	9.350	83	21085	2.06	ug/l #	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD031121\
Data File : VD068592.D
Acq On : 11 Mar 2021 18:30
Operator : VA/SY
Sample : M1642-03
Misc : 6.36G/5.00ml/MSVOA_D/SOIL
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
CSB-6-8-8.5

Quant Time: Mar 12 00:14:33 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D022621S.M
Quant Title : SW846 8260
QLast Update : Fri Feb 26 14:04:14 2021
Response via : Initial Calibration

