



Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD031122\
 Data File : VD072142.D
 Acq On : 11 Mar 2022 09:39
 Operator : VA/SY
 Sample : BFB
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D031122S.M
 Title : SW846 8260
 Last Update : Mon Mar 14 05:18:06 2022



AutoFind: Scans 1819, 1820, 1821; Background Corrected with Scan 1809

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	18.8	47344	PASS
75	95	30	60	50.4	126624	PASS
95	95	100	100	100.0	251456	PASS
96	95	5	9	6.6	16640	PASS
173	174	0.00	2	1.5	2880	PASS
174	95	50	100	78.8	198187	PASS
175	174	5	9	7.8	15446	PASS
176	174	95	101	97.0	192235	PASS
177	176	5	9	6.5	12452	PASS