

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD031221\
 Data File : VD068597.D
 Acq On : 12 Mar 2021 10:36
 Operator : VA/SY
 Sample : VD0312SBL01
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VD0312SBL01

Quant Time: Mar 12 23:33:32 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D022621S.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 26 14:04:14 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	421527	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	735456	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.643	117	717326	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.573	152	386063	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	208824	49.58	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	99.16%
35) Dibromofluoromethane	7.914	113	224051	50.57	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	101.14%
50) Toluene-d8	10.338	98	916423	51.93	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	103.86%
62) 4-Bromofluorobenzene	12.632	95	320328	56.32	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	112.64%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD031221\
Data File : VD068597.D
Acq On : 12 Mar 2021 10:36
Operator : VA/SY
Sample : VD0312SBL01
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VD0312SBL01

Quant Time: Mar 12 23:33:32 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D022621S.M
Quant Title : SW846 8260
QLast Update : Fri Feb 26 14:04:14 2021
Response via : Initial Calibration

