

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD031221\
 Data File : VD068604.D
 Acq On : 12 Mar 2021 14:08
 Operator : VA/SY
 Sample : M1625-04
 Misc : 6.39G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SP-1-2

Quant Time: Mar 12 23:35:30 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D022621S.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 26 14:04:14 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.985	168	333135	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.867	114	599566	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.644	117	601047	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.573	152	328750	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	181067	54.40	ug/l	0.00
Spiked Amount	50.000	Range 50 - 163	Recovery	= 108.80%		
35) Dibromofluoromethane	7.920	113	191757	53.09	ug/l	0.00
Spiked Amount	50.000	Range 54 - 147	Recovery	= 106.18%		
50) Toluene-d8	10.344	98	752580	52.31	ug/l	0.00
Spiked Amount	50.000	Range 49 - 140	Recovery	= 104.62%		
62) 4-Bromofluorobenzene	12.632	95	268047	57.81	ug/l	0.00
Spiked Amount	50.000	Range 25 - 144	Recovery	= 115.62%		
Target Compounds						Qvalue
16) Acetone	4.162	43	6701	12.67	ug/l	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD031221\
Data File : VD068604.D
Acq On : 12 Mar 2021 14:08
Operator : VA/SY
Sample : M1625-04
Misc : 6.39G/5.00ml/MSVOA_D/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SP-1-2

Quant Time: Mar 12 23:35:30 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D022621S.M
Quant Title : SW846 8260
QLast Update : Fri Feb 26 14:04:14 2021
Response via : Initial Calibration

