

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD031221\
 Data File : VD068606.D
 Acq On : 12 Mar 2021 15:05
 Operator : VA/SY
 Sample : M1649-05
 Misc : 5.82G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SP-1-3

Quant Time: Mar 12 23:35:56 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D022621S.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 26 14:04:14 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	314038	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.867	114	574761	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.644	117	556448	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.573	152	306064	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.338	65	165985	52.90	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 105.80%	
35) Dibromofluoromethane	7.914	113	179248	51.77	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 103.54%	
50) Toluene-d8	10.344	98	723850	52.49	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 104.98%	
62) 4-Bromofluorobenzene	12.632	95	249424	56.12	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 112.24%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD031221\
Data File : VD068606.D
Acq On : 12 Mar 2021 15:05
Operator : VA/SY
Sample : M1649-05
Misc : 5.82G/5.00ml/MSVOA_D/SOIL
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SP-1-3

Quant Time: Mar 12 23:35:56 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D022621S.M
Quant Title : SW846 8260
QLast Update : Fri Feb 26 14:04:14 2021
Response via : Initial Calibration

