

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD031221\
 Data File : VD068608.D
 Acq On : 12 Mar 2021 16:01
 Operator : VA/SY
 Sample : M1673-01
 Misc : 4.74G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 TT-040-RW6-IDWSO-20210311

Quant Time: Mar 12 23:36:23 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D022621S.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 26 14:04:14 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.985	168	132421	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.867	114	232008	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.644	117	238004	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.573	152	124094	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	75584	57.13	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	114.26%
35) Dibromofluoromethane	7.914	113	78101	55.88	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	111.76%
50) Toluene-d8	10.344	98	300048	53.90	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	107.80%
62) 4-Bromofluorobenzene	12.632	95	101134	56.37	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	112.74%

Target Compounds	Qvalue
-----	-----

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD031221\
Data File : VD068608.D
Acq On : 12 Mar 2021 16:01
Operator : VA/SY
Sample : M1673-01
Misc : 4.74G/5.00ml/MSVOA_D/SOIL
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
TT-040-RW6-IDWSO-20210311

Quant Time: Mar 12 23:36:23 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D022621S.M
Quant Title : SW846 8260
QLast Update : Fri Feb 26 14:04:14 2021
Response via : Initial Calibration

