



Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD031224\
Data File : VD078632.D
Acq On : 12 Mar 2024 08:41
Operator : RP/MD
Sample : BFB
Misc : 5.00G/5ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D030124S.M
Title : SW846 8260
Last Update : Sat Mar 02 03:50:17 2024



AutoFind: Scans 1867, 1868, 1869; Background Corrected with Scan 1859

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	18.8	12408	PASS
75	95	30	60	53.5	35283	PASS
95	95	100	100	100.0	65893	PASS
96	95	5	9	6.8	4500	PASS
173	174	0.00	2	1.5	921	PASS
174	95	50	100	94.5	62285	PASS
175	174	5	9	7.8	4875	PASS
176	174	95	101	97.2	60565	PASS
177	176	5	9	6.5	3947	PASS