



Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD031323\
 Data File : VD075503.D
 Acq On : 13 Mar 2023 10:07
 Operator : KP/SY
 Sample : BFB
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D031023S.M
 Title : SW846 8260
 Last Update : Fri Mar 10 23:52:37 2023



AutoFind: Scans 1867, 1868, 1869; Background Corrected with Scan 1859

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	15.8	9896	PASS
75	95	30	60	49.5	30888	PASS
95	95	100	100	100.0	62445	PASS
96	95	5	9	6.6	4110	PASS
173	174	0.00	2	1.2	668	PASS
174	95	50	100	89.8	56067	PASS
175	174	5	9	7.2	4055	PASS
176	174	95	101	98.9	55461	PASS
177	176	5	9	6.8	3769	PASS