

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD031419\
 Data File : VD061076.D
 Acq On : 14 Mar 2019 20:31
 Operator : VA/AP
 Sample : K1908-03
 Misc : 5.84g/5ml/MSVOA D/SOIL
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 101-102

Quant Time: Mar 15 03:52:44 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D030719S.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 08 02:35:39 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.06	168	522103	50.00	ug/l	0.01
34) 1,4-Difluorobenzene	8.23	114	925632	50.00	ug/l	0.01
63) Chlorobenzene-d5	12.37	117	714701	50.00	ug/l	0.01
72) 1,4-Dichlorobenzene-d4	14.52	152	274710	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.46	65	270815	62.32	ug/l	0.01
Spiked Amount 50.000			Recovery	=	124.64%	
35) Dibromofluoromethane	6.94	113	420908	61.27	ug/l	0.01
Spiked Amount 50.000			Recovery	=	122.54%	
50) Toluene-d8	10.41	98	918135	53.22	ug/l	0.01
Spiked Amount 50.000			Recovery	=	106.44%	
62) 4-Bromofluorobenzene	13.55	95	323774	48.37	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.74%	
Target Compounds						
16) Acetone	3.23	43	52655	42.370	ug/l	97
20) Methylene Chloride	3.84	84	50749	4.734	ug/l	93
52) Toluene	10.52	92	17852	1.440	ug/l	98
68) m/p-Xylenes	12.67	106	12095	1.477	ug/l	86
69) o-Xylene	13.05	106	9505	1.181	ug/l	87
80) 1,3,5-Trimethylbenzene	14.24	105	30585	2.035	ug/l	99
84) 1,2,4-Trimethylbenzene	14.24	105	30585	2.035	ug/l	99
95) Naphthalene	16.12	128	84793	15.988	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD031419\
Data File : VD061076.D
Acq On : 14 Mar 2019 20:31
Operator : VA/AP
Sample : K1908-03
Misc : 5.84g/5ml/MSVOA D/SOIL
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
101-102

Quant Time: Mar 15 03:52:44 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D030719S.M
Quant Title : SW846 8260
QLast Update : Fri Mar 08 02:35:39 2019
Response via : Initial Calibration

