

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD031419\
 Data File : VD061077.D
 Acq On : 14 Mar 2019 20:58
 Operator : VA/AP
 Sample : K1908-05
 Misc : 5.09g/5ml/MSVOA D/SOIL
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 93-94

Quant Time: Mar 15 03:54:20 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D030719S.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 08 02:35:39 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.06	168	580792	50.00	ug/l	0.02
34) 1,4-Difluorobenzene	8.23	114	1067481	50.00	ug/l	0.02
63) Chlorobenzene-d5	12.37	117	857260	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	14.52	152	313232	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.47	65	234142	48.43	ug/l	0.02
Spiked Amount 50.000			Recovery	=	96.86%	
35) Dibromofluoromethane	6.94	113	408090	51.51	ug/l	0.02
Spiked Amount 50.000			Recovery	=	103.02%	
50) Toluene-d8	10.42	98	959846	48.25	ug/l	0.02
Spiked Amount 50.000			Recovery	=	96.50%	
62) 4-Bromofluorobenzene	13.56	95	326259	42.27	ug/l	0.00
Spiked Amount 50.000			Recovery	=	84.54%	
Target Compounds						
16) Acetone	3.23	43	61296	44.339	ug/l	Qvalue 100
20) Methylene Chloride	3.83	84	54227	4.344	ug/l	97
25) 2-Butanone	6.08	43	8655	5.517	ug/l	95
52) Toluene	10.52	92	41486	2.902	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD031419\
Data File : VD061077.D
Acq On : 14 Mar 2019 20:58
Operator : VA/AP
Sample : K1908-05
Misc : 5.09g/5ml/MSVOA D/SOIL
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
93-94

Quant Time: Mar 15 03:54:20 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D030719S.M
Quant Title : SW846 8260
QLast Update : Fri Mar 08 02:35:39 2019
Response via : Initial Calibration

