

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD031422\
 Data File : VD072167.D
 Acq On : 14 Mar 2022 14:37
 Operator : VA/SY
 Sample : N1851-03
 Misc : 5.03G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 DUM-35

Quant Time: Mar 15 01:23:50 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D031122S.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 14 05:18:06 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	2622	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.867	114	4471	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	4604	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	1822	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.314	65	2504	85.693	ug/l	-0.01
Spiked Amount	50.000	Range	50 - 163	Recovery	=	171.380%#
35) Dibromofluoromethane	7.908	113	2253	71.608	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	143.220%
50) Toluene-d8	10.338	98	5608	48.747	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	97.500%
62) 4-Bromofluorobenzene	12.626	95	1855	41.498	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	83.000%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD031422\
Data File : VD072167.D
Acq On : 14 Mar 2022 14:37
Operator : VA/SY
Sample : N1851-03
Misc : 5.03G/5.00ml/MSVOA_D/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
DUM-35

Quant Time: Mar 15 01:23:50 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D031122S.M
Quant Title : SW846 8260
QLast Update : Mon Mar 14 05:18:06 2022
Response via : Initial Calibration

