

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD031422\
 Data File : VD072170.D
 Acq On : 14 Mar 2022 16:30
 Operator : VA/SY
 Sample : N1851-09
 Misc : 5.06G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 DUM-38

Quant Time: Mar 15 01:24:21 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D031122S.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 14 05:18:06 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	1548	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.850	114	2564	50.000	ug/l	-0.01
63) Chlorobenzene-d5	11.638	117	2650	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	703	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	1251	72.515	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	145.040%
35) Dibromofluoromethane	7.908	113	1367	75.763	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	151.520%#
50) Toluene-d8	10.338	98	3232	48.988	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	97.980%
62) 4-Bromofluorobenzene	12.632	95	732	28.555	ug/l	0.01
Spiked Amount	50.000	Range	25 - 144	Recovery	=	57.100%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD031422\
Data File : VD072170.D
Acq On : 14 Mar 2022 16:30
Operator : VA/SY
Sample : N1851-09
Misc : 5.06G/5.00ml/MSVOA_D/SOIL
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
DUM-38

Quant Time: Mar 15 01:24:21 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D031122S.M
Quant Title : SW846 8260
QLast Update : Mon Mar 14 05:18:06 2022
Response via : Initial Calibration

