

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_D\DATA\VD031519\
 Data File : VD061083.D
 Acq On : 15 Mar 2019 13:26
 Operator : VA/AP
 Sample : K1901-01RE
 Misc : 4.24g/5ml/MSVOA_D/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SB-1(0.5-1.5)RE

Quant Time: Mar 16 05:57:08 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D030719S.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 08 02:35:39 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.07	168	54961	50.00	ug/l	0.03
34) 1,4-Difluorobenzene	8.24	114	109019	50.00	ug/l	0.03
63) Chlorobenzene-d5	12.39	117	78064	50.00	ug/l	0.03
72) 1,4-Dichlorobenzene-d4	14.53	152	23627	50.00	ug/l	0.02
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.48	65	44125	96.45	ug/l	0.04
Spiked Amount 50.000			Recovery	=	192.90%	
35) Dibromofluoromethane	6.95	113	51530	63.69	ug/l	0.03
Spiked Amount 50.000			Recovery	=	127.38%	
50) Toluene-d8	10.43	98	95161	46.84	ug/l	0.03
Spiked Amount 50.000			Recovery	=	93.68%	
62) 4-Bromofluorobenzene	13.57	95	31115	39.47	ug/l	0.02
Spiked Amount 50.000			Recovery	=	78.94%	
Target Compounds						
20) Methylene Chloride	3.85	84	5909	5.782	ug/l #	78
64) Tetrachloroethene	11.27	164	20741	39.359	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_D\DATA\VD031519\
Data File : VD061083.D
Acq On : 15 Mar 2019 13:26
Operator : VA/AP
Sample : K1901-01RE
Misc : 4.24g/5ml/MSVOA_D/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SB-1(0.5-1.5)RE

Quant Time: Mar 16 05:57:08 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D030719S.M
Quant Title : SW846 8260
QLast Update : Fri Mar 08 02:35:39 2019
Response via : Initial Calibration

