

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_D\DATA\VD031519\
 Data File : VD061088.D
 Acq On : 15 Mar 2019 15:46
 Operator : VA/AP
 Sample : K1901-06RE
 Misc : 3.00g/5ml/MSVOA_D/SOIL
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SB-4(7-8)RE

Quant Time: Mar 16 06:25:38 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D030719S.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 08 02:35:39 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.08	168	321497	50.00	ug/l	0.04
34) 1,4-Difluorobenzene	8.25	114	615045	50.00	ug/l	0.04
63) Chlorobenzene-d5	12.38	117	322535	50.00	ug/l	0.03
72) 1,4-Dichlorobenzene-d4	14.53	152	58280	50.00	ug/l	0.02
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.48	65	206771	77.27	ug/l	0.04
Spiked Amount 50.000			Recovery	=	154.54%	
35) Dibromofluoromethane	6.96	113	302208	66.21	ug/l	0.04
Spiked Amount 50.000			Recovery	=	132.42%	
50) Toluene-d8	10.43	98	430341	37.54	ug/l	0.04
Spiked Amount 50.000			Recovery	=	75.08%	
62) 4-Bromofluorobenzene	13.57	95	97801	21.99	ug/l	0.02
Spiked Amount 50.000			Recovery	=	43.98%	
Target Compounds						
20) Methylene Chloride	3.85	84	45463	9.226	ug/l	99
64) Tetrachloroethene	11.29	164	3709	1.704	ug/l #	63

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_D\DATA\VD031519\
Data File : VD061088.D
Acq On : 15 Mar 2019 15:46
Operator : VA/AP
Sample : K1901-06RE
Misc : 3.00g/5ml/MSVOA_D/SOIL
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SB-4(7-8)RE

Quant Time: Mar 16 06:25:38 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D030719S.M
Quant Title : SW846 8260
QLast Update : Fri Mar 08 02:35:39 2019
Response via : Initial Calibration

