

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_D\DATA\VD031519\
 Data File : VD061093.D
 Acq On : 15 Mar 2019 18:10
 Operator : VA/AP
 Sample : K1908-01
 Misc : 5.69g/5ml/MSVOA_D/SOIL
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 MH-02

Quant Time: Mar 16 06:32:04 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D030719S.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 08 02:35:39 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.07	168	467940	50.00	ug/l	0.03
34) 1,4-Difluorobenzene	8.24	114	844442	50.00	ug/l	0.03
63) Chlorobenzene-d5	12.38	117	637285	50.00	ug/l	0.02
72) 1,4-Dichlorobenzene-d4	14.53	152	258350	50.00	ug/l	0.02
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.48	65	214053	54.96	ug/l	0.03
Spiked Amount 50.000			Recovery	=	109.92%	
35) Dibromofluoromethane	6.94	113	307660	49.09	ug/l	0.02
Spiked Amount 50.000			Recovery	=	98.18%	
50) Toluene-d8	10.43	98	705703	44.84	ug/l	0.03
Spiked Amount 50.000			Recovery	=	89.68%	
62) 4-Bromofluorobenzene	13.57	95	266085	43.57	ug/l	0.02
Spiked Amount 50.000			Recovery	=	87.14%	
Target Compounds						
16) Acetone	3.22	43	44615	40.056	ug/l	96
20) Methylene Chloride	3.84	84	37590	3.020	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_D\DATA\VD031519\
Data File : VD061093.D
Acq On : 15 Mar 2019 18:10
Operator : VA/AP
Sample : K1908-01
Misc : 5.69g/5ml/MSVOA_D/SOIL
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
MH-02

Quant Time: Mar 16 06:32:04 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D030719S.M
Quant Title : SW846 8260
QLast Update : Fri Mar 08 02:35:39 2019
Response via : Initial Calibration

