

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD031522\
 Data File : VD072191.D
 Acq On : 15 Mar 2022 16:18
 Operator : VA/SY
 Sample : N1905-03
 Misc : 3.70G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 WC-14A-1A-3-(3.0-3.5)

Quant Time: Mar 16 01:00:28 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D031122S.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 14 05:18:06 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	411223	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	798046	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	854682	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	401707	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	256327	55.932	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	111.860%
35) Dibromofluoromethane	7.908	113	259333	46.178	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	92.360%
50) Toluene-d8	10.338	98	1054382	51.346	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	102.700%
62) 4-Bromofluorobenzene	12.626	95	406756	50.979	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	101.960%
Target Compounds						
						Qvalue
16) Acetone	4.156	43	61261	61.567	ug/l	93
17) Carbon Disulfide	4.414	76	14753	3.124	ug/l #	89
25) 2-Butanone	7.202	43	20143	15.183	ug/l	96
31) Cyclohexane	7.979	56	114556	15.999	ug/l #	89
39) Methylcyclohexane	9.349	83	177177	19.985	ug/l	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD031522\
Data File : VD072191.D
Acq On : 15 Mar 2022 16:18
Operator : VA/SY
Sample : N1905-03
Misc : 3.70G/5.00ml/MSVOA_D/SOIL
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
WC-14A-1A-3-(3.0-3.5)

Quant Time: Mar 16 01:00:28 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D031122S.M
Quant Title : SW846 8260
QLast Update : Mon Mar 14 05:18:06 2022
Response via : Initial Calibration

