

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD031522\
 Data File : VD072200.D
 Acq On : 15 Mar 2022 20:34
 Operator : VA/SY
 Sample : N1912-01
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VNJ-242

Quant Time: Mar 16 01:03:38 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D031122S.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 14 05:18:06 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	62118	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	109361	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	128969	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	66144	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	43391	62.679	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 125.360%	
35) Dibromofluoromethane	7.908	113	37090	48.195	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 96.400%	
50) Toluene-d8	10.332	98	146720	52.140	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 104.280%	
62) 4-Bromofluorobenzene	12.620	95	66888	61.175	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 122.360%	
Target Compounds						
					Qvalue	
39) Methylcyclohexane	9.355	83	12995	10.696	ug/l	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD031522\
Data File : VD072200.D
Acq On : 15 Mar 2022 20:34
Operator : VA/SY
Sample : N1912-01
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VNJ-242

Quant Time: Mar 16 01:03:38 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D031122S.M
Quant Title : SW846 8260
QLast Update : Mon Mar 14 05:18:06 2022
Response via : Initial Calibration

