

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD031522\
 Data File : VD072209.D
 Acq On : 16 Mar 2022 01:44
 Operator : VA/SY
 Sample : N1922-02
 Misc : 5.29G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 7PH-2-8.5-9

Quant Time: Mar 16 03:45:11 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D031122S.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 14 05:18:06 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	639512	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	1091893	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	1163298	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	594832	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	320831	45.016	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	90.040%
35) Dibromofluoromethane	7.914	113	328864	42.800	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	85.600%
50) Toluene-d8	10.332	98	1356818	48.293	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	96.580%
62) 4-Bromofluorobenzene	12.620	95	591791	54.210	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	108.420%
Target Compounds						
17) Carbon Disulfide	4.409	76	5919	1.919	ug/l	Qvalue # 92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD031522\
Data File : VD072209.D
Acq On : 16 Mar 2022 01:44
Operator : VA/SY
Sample : N1922-02
Misc : 5.29G/5.00ml/MSVOA_D/SOIL
ALS Vial : 34 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
7PH-2-8.5-9

Quant Time: Mar 16 03:45:11 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D031122S.M
Quant Title : SW846 8260
QLast Update : Mon Mar 14 05:18:06 2022
Response via : Initial Calibration

