

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD031524\
 Data File : VD078684.D
 Acq On : 15 Mar 2024 15:47
 Operator : RP/MD
 Sample : P1746-15RE
 Misc : 5.02G/5ml/MSVOA_D/SOIL/B
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SB-05-0-2.0RE

Quant Time: Mar 16 05:26:26 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D031424S.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 15 05:38:03 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.869	168	1454	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.775	114	3374	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.587	117	3835	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	1437	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	2066	151.618	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 303.240%#	
35) Dibromofluoromethane	7.793	113	1939	89.258	ug/l	-0.01
Spiked Amount	50.000	Range	54 - 147	Recovery	= 178.520%#	
50) Toluene-d8	10.269	98	4071	50.956	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 101.920%	
62) 4-Bromofluorobenzene	12.575	95	1282	54.850	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	= 109.700%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD031524\
Data File : VD078684.D
Acq On : 15 Mar 2024 15:47
Operator : RP/MD
Sample : P1746-15RE
Misc : 5.02G/5ml/MSVOA_D/SOIL/B
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SB-05-0-2.0RE

Quant Time: Mar 16 05:26:26 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D031424S.M
Quant Title : SW846 8260
QLast Update : Fri Mar 15 05:38:03 2024
Response via : Initial Calibration

