

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD031524\
 Data File : VD078685.D
 Acq On : 15 Mar 2024 16:14
 Operator : RP/MD
 Sample : P1746-16RE
 Misc : 5.04G/5ml/MSVOA_D/SOIL/B
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SB-05-0-2.0-DUPRE

Quant Time: Mar 16 05:26:47 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D031424S.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 15 05:38:03 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	61891	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	111671	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	88732	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	22571	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.234	65	43179	74.444	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	148.880%
35) Dibromofluoromethane	7.810	113	46802	65.094	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	130.180%
50) Toluene-d8	10.269	98	132796	50.221	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	100.440%
62) 4-Bromofluorobenzene	12.569	95	23514	30.396	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	60.800%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD031524\
Data File : VD078685.D
Acq On : 15 Mar 2024 16:14
Operator : RP/MD
Sample : P1746-16RE
Misc : 5.04G/5ml/MSVOA_D/SOIL/B
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SB-05-0-2.0-DUPRE

Quant Time: Mar 16 05:26:47 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D031424S.M
Quant Title : SW846 8260
QLast Update : Fri Mar 15 05:38:03 2024
Response via : Initial Calibration

