

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD031524\
 Data File : VD078686.D
 Acq On : 15 Mar 2024 16:42
 Operator : RP/MD
 Sample : P1746-01RE
 Misc : 5.08G/5ml/MSVOA_D/SOIL/B
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SB-01-0-2.0RE

Quant Time: Mar 16 05:27:07 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D031424S.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 15 05:38:03 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.881	168	72383	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	132702	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	116120	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	37788	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.234	65	45528	67.116	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	134.240%
35) Dibromofluoromethane	7.810	113	49494	57.928	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	115.860%
50) Toluene-d8	10.269	98	161114	51.274	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	102.540%
62) 4-Bromofluorobenzene	12.575	95	35069	38.149	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	76.300%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD031524\
Data File : VD078686.D
Acq On : 15 Mar 2024 16:42
Operator : RP/MD
Sample : P1746-01RE
Misc : 5.08G/5ml/MSVOA_D/SOIL/B
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SB-01-0-2.0RE

Quant Time: Mar 16 05:27:07 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D031424S.M
Quant Title : SW846 8260
QLast Update : Fri Mar 15 05:38:03 2024
Response via : Initial Calibration

