


Data Path : Z:\voasrv\HPCHEM1\MSVOA D\Data\VD031619\
Data File : VD061101.D
Acq On : 16 Mar 2019 9:36
Operator : VA/AP
Sample : BFB
Misc : 5.00g/5ml/MSVOA D/SOIL
ALS Vial : 2 Sample Multiplier: 1

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D030719S.M
Title : SW846 8260
Last Update : Fri Mar 08 02:35:39 2019



AutoFind: Scans 1232, 1233, 1234; Background Corrected with Scan 1228

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	17.8	18806	PASS
75	95	30	60	37.1	39088	PASS
95	95	100	100	100.0	105498	PASS
96	95	5	9	6.6	6951	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	78.1	82426	PASS
175	174	5	9	7.2	5910	PASS
176	174	95	101	97.0	79946	PASS
177	176	5	9	6.6	5293	PASS