


Data Path : Z:\voasrv\HPCHEM1\MSVOA D\Data\VD031619\
Data File : VD061122.D
Acq On : 16 Mar 2019 22:03
Operator : VA/AP
Sample : BFB
Misc : 6.00q/5ml/MSVOA D/SOIL
ALS Vial : 25 Sample Multiplier: 1

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D030719S.M
Title : SW846 8260
Last Update : Fri Mar 08 02:35:39 2019



AutoFind: Scans 1231, 1232, 1233; Background Corrected with Scan 1227

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	17.6	17411	PASS
75	95	30	60	37.4	36952	PASS
95	95	100	100	100.0	98802	PASS
96	95	5	9	6.3	6194	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	76.7	75768	PASS
175	174	5	9	7.1	5386	PASS
176	174	95	101	96.5	73123	PASS
177	176	5	9	6.5	4779	PASS