

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD031621\
 Data File : VD068624.D
 Acq On : 16 Mar 2021 11:35
 Operator : VA/SY
 Sample : VD0316SBL01
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VD0316SBL01

Quant Time: Mar 17 01:04:24 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D022621S.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 26 14:04:14 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	346505	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.867	114	605554	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.643	117	583051	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.573	152	310934	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	166692	48.15	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	96.30%
35) Dibromofluoromethane	7.914	113	186540	51.14	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	102.28%
50) Toluene-d8	10.344	98	761477	52.41	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	104.82%
62) 4-Bromofluorobenzene	12.632	95	262811	56.12	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	112.24%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD031621\
Data File : VD068624.D
Acq On : 16 Mar 2021 11:35
Operator : VA/SY
Sample : VD0316SBL01
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VD0316SBL01

Quant Time: Mar 17 01:04:24 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D022621S.M
Quant Title : SW846 8260
QLast Update : Fri Feb 26 14:04:14 2021
Response via : Initial Calibration









Pentafluorobenzene
Concen: 50.00 ug/l
RT: 7.979 min Scan# 1030
Delta R.T. -0.006 min
Lab File: VD068624.D
Acq: 16 Mar 2021 11:35




Instrument :
MSVOA_D
ClientSampleId :
VD0316SBL01


Tgt Ion:168 Resp: 346505
Ion Ratio Lower Upper
168 100
99 54.7 43.5 65.3



1,2-Dichloroethane-d4
Concen: 48.15 ug/l
RT: 8.332 min Scan# 1090
Delta R.T. 0.000 min
Lab File: VD068624.D
Acq: 16 Mar 2021 11:35

Tgt Ion: 65 Resp: 166692
Ion Ratio Lower Upper
65 100
67 53.2 0.0 105.0









#34
1,4-Difluorobenzene
Concen: 50.00 ug/l
RT: 8.867 min Scan# 1181
Delta R.T. 0.000 min
Lab File: VD068624.D
Acq: 16 Mar 2021 11:35




Instrument :
MSVOA_D
ClientSampleId :
VD0316SBL01


Tgt Ion:114 Resp: 605554
Ion Ratio Lower Upper
114 100
63 18.9 0.0 34.8
88 14.4 0.0 31.6



#35
Dibromofluoromethane
Concen: 51.14 ug/l
RT: 7.914 min Scan# 1019
Delta R.T. -0.006 min
Lab File: VD068624.D
Acq: 16 Mar 2021 11:35

Tgt Ion:113 Resp: 186540
Ion Ratio Lower Upper
113 100
111 101.7 81.4 122.0
192 18.2 16.1 24.1









#50
Toluene-d8
Concen: 52.41 ug/l
RT: 10.344 min Scan# 1432
Delta R.T. 0.000 min
Lab File: VD068624.D
Acq: 16 Mar 2021 11:35




Instrument :
MSVOA_D
ClientSampleId :
VD0316SBL01


Tgt Ion: 98 Resp: 761477
Ion Ratio Lower Upper
98 100
100 64.6 53.0 79.4



#62
4-Bromofluorobenzene
Concen: 56.12 ug/l
RT: 12.632 min Scan# 1821
Delta R.T. 0.000 min
Lab File: VD068624.D
Acq: 16 Mar 2021 11:35

Tgt Ion: 95 Resp: 262811
Ion Ratio Lower Upper
95 100
174 85.2 0.0 172.4
176 80.5 0.0 168.4









#63
Chlorobenzene-d5
Concen: 50.00 ug/l
RT: 11.643 min Scan# 1653
Delta R.T. 0.000 min
Lab File: VD068624.D
Acq: 16 Mar 2021 11:35

Instrument :
MSVOA_D
ClientSampleId :
VD0316SBL01

Tgt Ion:	117	Resp:	583051
Ion	Ratio	Lower	Upper
117	100		
82	55.6	41.4	62.0
119	31.9	26.1	39.1



#72
1,4-Dichlorobenzene-d4
Concen: 50.00 ug/l
RT: 13.573 min Scan# 1981
Delta R.T. 0.001 min
Lab File: VD068624.D
Acq: 16 Mar 2021 11:35

Tgt Ion:	152	Resp:	310934
Ion	Ratio	Lower	Upper
152	100		
115	58.3	28.9	86.8
150	159.8	0.0	344.4

