

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD031622\
 Data File : VD072216.D
 Acq On : 16 Mar 2022 14:12
 Operator : VA/SY
 Sample : N1918-15
 Misc : 4.91G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 JDSH-TB-STOCKPILE-C

Quant Time: Mar 16 15:01:21 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D031122S.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 14 05:18:06 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	486975	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	800254	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.637	117	827992	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	403014	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	235896	43.467	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	86.940%
35) Dibromofluoromethane	7.908	113	239501	42.529	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	85.060%
50) Toluene-d8	10.332	98	975360	47.367	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	94.740%
62) 4-Bromofluorobenzene	12.626	95	405351	50.663	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	101.320%
Target Compounds						
17) Carbon Disulfide	4.414	76	17591	3.135	ug/l	Qvalue 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD031622\
Data File : VD072216.D
Acq On : 16 Mar 2022 14:12
Operator : VA/SY
Sample : N1918-15
Misc : 4.91G/5.00ml/MSVOA_D/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
JDSH-TB-STOCKPILE-C

Quant Time: Mar 16 15:01:21 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D031122S.M
Quant Title : SW846 8260
QLast Update : Mon Mar 14 05:18:06 2022
Response via : Initial Calibration

