

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD031622\
 Data File : VD072221.D
 Acq On : 16 Mar 2022 16:34
 Operator : VA/SY
 Sample : N1914-01
 Misc : 4.99G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 TP-1

Quant Time: Mar 16 16:46:12 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D031122S.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 14 05:18:06 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	540882	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	889145	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	958195	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	479448	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	246172	40.839	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	81.680%
35) Dibromofluoromethane	7.909	113	262403	41.938	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	83.880%
50) Toluene-d8	10.332	98	1118638	48.894	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	97.780%
62) 4-Bromofluorobenzene	12.620	95	472555	53.158	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	106.320%
Target Compounds						
16) Acetone	4.150	43	13957	10.664	ug/l	Qvalue 94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD031622\
Data File : VD072221.D
Acq On : 16 Mar 2022 16:34
Operator : VA/SY
Sample : N1914-01
Misc : 4.99G/5.00ml/MSVOA_D/SOIL
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
TP-1

Quant Time: Mar 16 16:46:12 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D031122S.M
Quant Title : SW846 8260
QLast Update : Mon Mar 14 05:18:06 2022
Response via : Initial Calibration

