

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD031622\
 Data File : VD072225.D
 Acq On : 16 Mar 2022 18:27
 Operator : VA/SY
 Sample : N1893-05
 Misc : 5.02G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 CONCRETE

Quant Time: Mar 17 02:28:13 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D031122S.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 14 05:18:06 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	402623	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	671850	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	738864	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	376322	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	215737	48.080	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	96.160%
35) Dibromofluoromethane	7.914	113	101461	21.460	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	42.920%#
50) Toluene-d8	10.332	98	847753	49.039	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	98.080%
62) 4-Bromofluorobenzene	12.620	95	375611	55.918	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	111.840%
Target Compounds						
16) Acetone	4.156	43	16059	16.484	ug/l	Qvalue 94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD031622\
Data File : VD072225.D
Acq On : 16 Mar 2022 18:27
Operator : VA/SY
Sample : N1893-05
Misc : 5.02G/5.00ml/MSVOA_D/SOIL
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
CONCRETE

Quant Time: Mar 17 02:28:13 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D031122S.M
Quant Title : SW846 8260
QLast Update : Mon Mar 14 05:18:06 2022
Response via : Initial Calibration

