

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD031622\
 Data File : VD072226.D
 Acq On : 16 Mar 2022 18:56
 Operator : VA/SY
 Sample : N1884-01
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 HR-03-031022

Quant Time: Mar 17 02:28:30 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D031122S.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 14 05:18:06 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	395342	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	662223	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	726967	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	371045	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	203057	46.088	ug/l	0.00
Spiked Amount 50.000	Range 50 - 163		Recovery =	92.180%		
35) Dibromofluoromethane	7.909	113	202574	43.470	ug/l	0.00
Spiked Amount 50.000	Range 54 - 147		Recovery =	86.940%		
50) Toluene-d8	10.332	98	831360	48.789	ug/l	0.00
Spiked Amount 50.000	Range 49 - 140		Recovery =	97.580%		
62) 4-Bromofluorobenzene	12.626	95	363961	54.972	ug/l	0.00
Spiked Amount 50.000	Range 25 - 144		Recovery =	109.940%		
Target Compounds						Qvalue
17) Carbon Disulfide	4.409	76	4388	2.003	ug/l #	81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD031622\
Data File : VD072226.D
Acq On : 16 Mar 2022 18:56
Operator : VA/SY
Sample : N1884-01
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
HR-03-031022

Quant Time: Mar 17 02:28:30 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D031122S.M
Quant Title : SW846 8260
QLast Update : Mon Mar 14 05:18:06 2022
Response via : Initial Calibration

