

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD031622\
 Data File : VD072228.D
 Acq On : 16 Mar 2022 19:53
 Operator : VA/SY
 Sample : N1912-01
 Misc : 5.06G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VNJ-242RE

Quant Time: Mar 17 02:28:56 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D031122S.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 14 05:18:06 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	30169	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.861	114	49040	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.637	117	58353	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	28148	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	23517	69.946	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	139.900%
35) Dibromofluoromethane	7.914	113	19126	55.422	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	110.840%
50) Toluene-d8	10.332	98	57930	45.909	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	91.820%
62) 4-Bromofluorobenzene	12.620	95	26844	54.750	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	109.500%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD031622\
Data File : VD072228.D
Acq On : 16 Mar 2022 19:53
Operator : VA/SY
Sample : N1912-01
Misc : 5.06G/5.00ml/MSVOA_D/SOIL
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VNJ-242RE

Quant Time: Mar 17 02:28:56 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D031122S.M
Quant Title : SW846 8260
QLast Update : Mon Mar 14 05:18:06 2022
Response via : Initial Calibration

