

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD031819\
 Data File : VD061158.D
 Acq On : 18 Mar 2019 21:09
 Operator : VA/AP
 Sample : K1935-10
 Misc : 4.80u/5ml/MSVOA D/SOIL
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SB-7(6-7)

Quant Time: Mar 19 04:40:37 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D030719S.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 08 02:35:39 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.03	168	99641	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.22	114	237637	50.00	ug/l	0.00
63) Chlorobenzene-d5	12.36	117	118099	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	14.52	152	25796	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.45	65	86762	104.61	ug/l	0.00
Spiked Amount 50.000			Recovery	=	209.22%	
35) Dibromofluoromethane	6.92	113	100660	57.08	ug/l	0.00
Spiked Amount 50.000			Recovery	=	114.16%	
50) Toluene-d8	10.40	98	128925	29.11	ug/l	0.00
Spiked Amount 50.000			Recovery	=	58.22%	
62) 4-Bromofluorobenzene	13.55	95	36243	21.09	ug/l	0.00
Spiked Amount 50.000			Recovery	=	42.18%	
Target Compounds						
16) Acetone	3.21	43	84366	355.715	ug/l	96
20) Methylene Chloride	3.81	84	7254	2.254	ug/l	92
25) 2-Butanone	6.06	43	26397	98.084	ug/l	95
39) Methylcyclohexane	8.85	83	28492	13.414	ug/l #	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD031819\
Data File : VD061158.D
Acq On : 18 Mar 2019 21:09
Operator : VA/AP
Sample : K1935-10
Misc : 4.80µl/5ml/MSVOA D/SOIL
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SB-7(6-7)

Quant Time: Mar 19 04:40:37 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D030719S.M
Quant Title : SW846 8260
QLast Update : Fri Mar 08 02:35:39 2019
Response via : Initial Calibration

