

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD031821\
 Data File : VD068656.D
 Acq On : 18 Mar 2021 16:06
 Operator : VA/SY
 Sample : M1686-01
 Misc : 5.01G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SW-VAN-WINKLE-STREET

Quant Time: Mar 19 01:06:09 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D022621S.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 26 14:04:14 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.985	168	77140	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.867	114	143302	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.643	117	152374	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.573	152	86580	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	57230	74.25	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	148.50%
35) Dibromofluoromethane	7.914	113	24572	28.46	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	56.92%
50) Toluene-d8	10.338	98	180076	52.37	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	104.74%
62) 4-Bromofluorobenzene	12.632	95	70484	63.61	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	127.22%
Target Compounds						
16) Acetone	4.162	43	6577	53.72	ug/l	Qvalue # 88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD031821\
Data File : VD068656.D
Acq On : 18 Mar 2021 16:06
Operator : VA/SY
Sample : M1686-01
Misc : 5.01G/5.00ml/MSVOA_D/SOIL
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SW-VAN-WINKLE-STREET

Quant Time: Mar 19 01:06:09 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D022621S.M
Quant Title : SW846 8260
QLast Update : Fri Feb 26 14:04:14 2021
Response via : Initial Calibration

